



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

Handwritten initials and date:
2/3/16

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-006048 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1-26-2016 Cal 1 — AC Cal 2 — AC

Between the dates of — and —
the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 1-26-2016.
The instrument was returned to service on —.

Handwritten signature of Terry S. Gallegos

Terry S. Gallegos

1-27-2016


ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 27th, day of January, 2016.

Handwritten signature of Elma Escarcega

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5





TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

1 of 2
TS4
JLS

Date: 1-26-2016

QAS: Terry Gallegos 28847

Intoxilyzer Serial Number: 80-006048

Reason For Service: (on 1-7-2016) I was notified by Officer Knepper that
80-006048 displayed "system fault #018 Contact CMI"
and I brought the Intoxilyzer to the Lab for more
tests. I downloaded the COBRA data and sent it to CMI
for repair. TS4

Date Out of Service: 1-26-2016

Calibration Check completed prior to removal from service

Date: No

SQAP completed prior to removal from service

Date: No

Corrective Action: ☒ Sent to CMI, Inc.

20f2
D4
DA

EMI INC.
Service Evaluation Form

This form MUST be completed and enclosed with instrument to be serviced.

Note: Please ship items in their original shipping container.

Contact information: Customer Number 105075 (contact Customer Service)
Name TERRY GALLEGOS Phone: (520) 837-7677
Fax: (520) 791-4642 Email: TERRY.GALLEGOS@TUCSONAZ.GOV

Bill to Address:	Ship to Address:
<u>FINANCE/ACCOUNTS PAYABLE</u>	<u>TUCSON POLICE DEPARTMENT CRIMELAB</u>
<u>PO BOX 27450</u>	<u>1306 W. MIRAGE MILE</u>
<u>TUCSON, AZ 85726-7450</u>	<u>TUCSON, AZ 85705</u>

Instrument Serial Number: 80-006048

Detailed Description of Problem:

Display "System Fault #018 Contact CMI" and it failed these
Diagnostics: EEPROM, DSP, Analytical Test.

***Hazardous Material Warning! - DO NOT return gas cylinders with instrument! ***

Note: An evaluation fee of \$39.50 for handheld instruments or \$79.00 for IR evidential instruments will apply to estimates which are not repaired.

☐ I Authorize Repairs Up To: ☐ All ☐ \$250 ☐ \$500 ☐ \$750 ☐ Other \$ _____
Purchase Order Number 37768

Authorized By:

TERRY GALLEGOS TOXICOLOGY COORDINATOR

Name (Please Print)

Title

Terry Gallegos

Signature

1.26.2016

Date

Ship to:

CMI, Inc.
Attn: Service Dept.
316 East Ninth Street
Owensboro, KY 42303

☒ No, please send estimate before repairs are made. An estimate will be faxed before performing any repairs and may cause delays in service.